

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address COMMISSIONER FOR PATENTS FO Box 1430 Alexandra, Virginia 22313-1450 www.tepto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/658,116	09/08/2003	John P. Willis	BVTP-P02-004	9687
28120 7879 9221/2998 ROPES & GRAY ILP PATENT DOCKETING 39/41 ONE INTERNATIONAL PLACE BOSTON, MA 02110-2624			EXAMINER	
			STIGELL, THEODORE J	
			ART UNIT	PAPER NUMBER
2001011,111	02110 2021		3763	
			MAIL DATE	DELIVERY MODE
			02/21/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

# Application No. Applicant(s) 10/658,116 WILLIS ET AL. Office Action Summary Examiner Art Unit THEODORE J. STIGELL 3763 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS. WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status 1) Responsive to communication(s) filed on 05 November 2007. 2a) This action is FINAL. 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. Disposition of Claims 4) Claim(s) 1 and 27-35 is/are pending in the application. 4a) Of the above claim(s) is/are withdrawn from consideration. 5) Claim(s) \_\_\_\_\_ is/are allowed. 6) Claim(s) 1.27-33 and 35 is/are rejected. 7) Claim(s) 34 is/are objected to. 8) Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement. Application Papers 9) The specification is objected to by the Examiner. 10) ☐ The drawing(s) filed on is/are: a) ☐ accepted or b) ☐ objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abevance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. Priority under 35 U.S.C. § 119 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some \* c) None of: Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). \* See the attached detailed Office action for a list of the certified copies not received.

Application/Control Number: 10/658,116

Art Unit: 3763

#### DETAILED ACTION

## Response to Amendment

#### Specification

The amendment to the specification filed 11/5/2007 is acknowledged and accepted.

### Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

- (b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.
- (e) the invention was described in (1) an application for patent, published under section 122(b), by another filled in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 29-33 and 35 are rejected under 35 U.S.C. 102(b) as being anticipated by Lindmayer et al. (4,623,332). Lindmayer discloses an injection device (See Figures 6-8) comprising at least a first housing (45), a second housing (2), a third housing (5), and a fourth housing (3), wherein the third housing has a syringe adaptor (threaded portion of 5, the limitation of "syringe adaptor" does not carry much patentable weight by itself) having an outer vent sleeve (58, the examiner is interpreting a "vent sleeve" to basically mean just a sleeve that has some venting capability which any sleeve does), a movable piston (51) having an end and at least one relief hole (the bifurcation), a fixed sleeve (not numbered, element that is threaded to the threads of element 5) adjacent the

Application/Control Number: 10/658,116

Art Unit: 3763

movable piston and having a groove (threads) with a hole (the element is a cylinder and therefore has a hole through it), a drive piston (61) having a groove and a sealing device (54) within the groove, wherein the housings are demountably attached, wherein the first housing has a button (49) and the fourth housing has a plunger (47), a syringe, and a spring (50).

Claims 1 and 27-28 are rejected under 35 U.S.C. 102(e) as being anticipated by Jacobsen et al. (6,165,155). See Figures 8A-8D and the respective portions of the specification. Jacobsen discloses an injection device comprising a housing (500) having a proximal end and a distal end, the housing having a distal opening (538), a propellant (526, 528) disposed inside the housing and spaced from the distal end, a piston (510) having a first cavity disposed inside the housing between the distal end and the propellant, and a hollow sleeve (512) configured to mate with the piston, the sleeve having a second cavity, wherein a gas generated by the propellant\_travels between the interior of the housing to the exterior of the injection device through the first cavity and the second cavity but without passing through the distal opening of the housing, wherein the hollow sleeve has a surface with at least one hole (544), and wherein the piston has at least one hole (546) and a surface adjacent the surface of the sleeve.

#### Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., In re Berg, 140

Application/Control Number: 10/658,116

Art Unit: 3763

F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Omum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1 and 27-35 are rejected on the ground of n0nstatutory obviousness-type double patenting as being unpatentable over claims 1-23 of U.S. Patent No. 6,616,627. Although the conflicting claims are not identical, they are not patentably distinct from each other because the minor structural differences in the sets of claims do not pass the threshold of patentability.

### Allowable Subject Matter

Claim 34 is objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

#### Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

Application/Control Number: 10/658,116
Art Unit: 3763

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to THEODORE J. STIGELL whose telephone number is (571)272-8759. The examiner can normally be reached on M-F 8:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/658,116 Page 6

Art Unit: 3763

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Theodore J Stigell/ Examiner, Art Unit 3763

/Nicholas D Lucchesi/

Supervisory Patent Examiner, Art Unit 3763